
FDA-Konformitätserklärung

Declaration of conformity

Als Hersteller **Seli GmbH Automatisierungstechnik**
Dieselstraße 13
D-48485 Neuenkirchen

Geräteart: Elektronischer Temperatursensor

Typ: TF53, TF55

1. Aseptoflex Dichkonzept 2. Abdichtung über 1/2 Dichtkonus 3. Prozessadaption

dass die mit Medium in Kontakt stehenden Oberflächen bzw. Materialien, bestehend aus Edelstahl 316 L (f1.4404 / 1.4435) und / oder einem der folgenden Grundstoffe:
- Elastomer (FKM/EPDM), - Kunststoff (PEEK) u. a. auf Grundlage der Lieferantenbescheinigungen konform sind zu:
CFR Title 21 (FDA) § 177.2415

As manufacturer: **Seli GmbH Automatisierungstechnik**
Dieselstraße 13
D-48485 Neuenkirchen

Unit: Electronic Temperature Sensor

Type: TF53, TF55

1. Aseptoflex Vario sealing concept 2. Sealing via G1/2 sealing cone 3. Process adapters

that the surfaces / materials in contact with the medium, consisting of stainless steel 316L (1.4404 / 1.4435) and / or on of the following basic materials:
- elastomer (FKM / EPDM), - plastics (PEEK) are, among others on the basis of supplier declarations compliant with:

CFR Title 21 (FDA) § 177.2415

Neuenkirchen, 14. April 2016


Seli GmbH Automatisierungstechnik